

CSUS


by Allard USA

C|s

CLAVICULAR SPLINT
Rx Only — Single Patient Use Only

C|s

Instructions for Use



FEEL THE FUNCTION

The intended function of this splint is to offer support to stabilize the clavicle. This splint can also be used to improve posture. Review instructions on back page carefully before application.

For questions, please contact your physician, the professional who fit this device or Allard USA Customer Service:

CALL: 888-678-6548 FAX: 800-289-0809
info@allardusa.com

Allard USA, Inc. 300 Forge Way, Suite 3 Rockaway, NJ 07866-2056
www.allardusa.com

CSUS


by Allard USA

C|s

CLAVICULAR SPLINT
Rx Only — Single Patient Use Only

C|s

Instructions for Use



FEEL THE FUNCTION

The intended function of this splint is to offer support to stabilize the clavicle. This splint can also be used to improve posture. Review instructions on back page carefully before application.


For questions, please contact your physician, the professional who fit this device or Allard USA Customer Service:

CALL: 888-678-6548 FAX: 800-289-0809
info@allardusa.com


Allard USA, Inc. 300 Forge Way, Suite 3 Rockaway, NJ 07866-2056
www.allardusa.com

Clavicular Splint

1. Pull the tabs with hook loose from the padding and center the round pad with D-Rings, between the shoulder blades. Allow the padded straps to lay comfortably over each shoulder.



2. Pull the padded straps under the arms and feed the velcro tabs through the D-Rings as shown. Grasp both velcro tabs and pull simultaneously until shoulders are brought back to desired position. Secure in position by pressing the tabs with hook into the padding.



Indications / Intended Use

Clavicle fractures, grade II & III acromioclavicular separations, osteolysis of the clavicle, AC degenerative joint pain, mechanical cervical disease, postural dysfunction, sternoclavicular dislocations.

Contraindications

Osteopenia, osteoporosis.

Wash instruction

Hand wash in cold water using mild detergent.
Air dry.

Adjustment


This splint should only be fitted by a licensed medical professional.

Warning


Do not overtighten, as this can cause poor circulation and/or skin irritation. If you feel any irritations or other problems, remove splint. Check skin on a daily basis.

Clavicular Splint

1. Pull the tabs with hook loose from the padding and center the round pad with D-Rings, between the shoulder blades. Allow the padded straps to lay comfortably over each shoulder.



2. Pull the padded straps under the arms and feed the velcro tabs through the D-Rings as shown. Grasp both velcro tabs and pull simultaneously until shoulders are brought back to desired position. Secure in position by pressing the tabs with hook into the padding.



Indications / Intended Use

Clavicle fractures, grade II & III acromioclavicular separations, osteolysis of the clavicle, AC degenerative joint pain, mechanical cervical disease, postural dysfunction, sternoclavicular dislocations.

Contraindications

Osteopenia, osteoporosis.

Wash instruction

Hand wash in cold water using mild detergent.
Air dry.

Adjustment

This splint should only be fitted by a licensed medical professional.

Warning

Do not overtighten, as this can cause poor circulation and/or skin irritation. If you feel any irritations or other problems, remove splint. Check skin on a daily basis.