

# CSUS

by Allard USA

C|s

## ABDUCTION PILLOW with SLING

Rx Only — Single Patient Use Only

C|s

### Instructions for Use



FEEL THE FUNCTION

*The intended function of this sling is to transfer the weight of the forearm to the patient's back and unaffected shoulder while abducting the arm away from the body at about 10°-15° for shoulder rehabilitation.*

For questions, please contact your physician, the professional who fit this device or Allard USA Customer Service:

CALL: 888-678-6548      FAX: 800-289-0809  
info@allardusa.com

Allard USA, Inc. 300 Forge Way, Suite 3 Rockaway, NJ 07866-2056  
www.allardusa.com

# CSUS

by Allard USA

C|s

## ABDUCTION PILLOW with SLING

Rx Only — Single Patient Use Only

C|s

### Instructions for Use



FEEL THE FUNCTION

*The intended function of this sling is to transfer the weight of the forearm to the patient's back and unaffected shoulder while abducting the arm away from the body at about 10°-15° for shoulder rehabilitation.*

For questions, please contact your physician, the professional who fit this device or Allard USA Customer Service:

CALL: 888-678-6548      FAX: 800-289-0809  
info@allardusa.com

Allard USA, Inc. 300 Forge Way, Suite 3 Rockaway, NJ 07866-2056  
www.allardusa.com

## Abduction Pillow with Sling

1. Release all straps and buckles. Place affected arm in the fold of the sling, Make sure elbow is seated in corner.



2. Loop strap over unaffected shoulder and pass through both D-rings. Secure strap back against self and adjust to desired angle. Attach short strap to secure bicep.

3. Align foam pillow at waistline on affected side under sling as shown. Wrap strap around waist and buckle. Adjust strap length as needed for snug comfortable fit.



### Indications / Intended Use

Rotator cuff repairs; Anterior repairs; Posterior dislocations; Capsular shifts; Global shoulder instability

### Contraindications

Inability to bend arm to at least 90°.

### Wash instruction

Hand wash in cold water using mild detergent.  
Air dry.

### Adjustment

This sling should only be fitted by a licensed medical professional.


### Warning

If wearing the sling causes any unusual reactions, swelling or prolonged pain, remove and consult your medical professional. Check skin on a daily basis.

**CSUS**  
by Allard USA

## Abduction Pillow with Sling

1. Release all straps and buckles. Place affected arm in the fold of the sling, Make sure elbow is seated in corner.



2. Loop strap over unaffected shoulder and pass through both D-rings. Secure strap back against self and adjust to desired angle. Attach short strap to secure bicep.

3. Align foam pillow at waistline on affected side under sling as shown. Wrap strap around waist and buckle. Adjust strap length as needed for snug comfortable fit.



### Indications / Intended Use

Rotator cuff repairs; Anterior repairs; Posterior dislocations; Capsular shifts; Global shoulder instability

### Contraindications

Inability to bend arm to at least 90°.

### Wash instruction

Hand wash in cold water using mild detergent.  
Air dry.

### Adjustment

This sling should only be fitted by a licensed medical professional.

### Warning

If wearing the sling causes any unusual reactions, swelling or prolonged pain, remove and consult your medical professional. Check skin on a daily basis.

**CSUS**  
by Allard USA