

# CSUS

by Allard USA

C|s

## ROM HINGE KNEE ADJ THIGH

Rx Only — Single Patient Use Only

C|s

### Instructions for Use



FEEL THE FUNCTION

*The intended function of this sleeve is to provide range of motion control and medial/lateral stability through compression for the treatment of mild knee instabilities. Review instructions on back page carefully before application.*

For questions, please contact your physician, the professional who fit this device or Allard USA Customer Service:

CALL: 888-678-6548      FAX: 800-289-0809  
info@allardusa.com

Allard USA, Inc. 300 Forge Way, Suite 3 Rockaway, NJ 07866-2056  
www.allardusa.com

# CSUS

by Allard USA


C|s

## ROM HINGE KNEE ADJ THIGH

Rx Only — Single Patient Use Only

C|s

### Instructions for Use



FEEL THE FUNCTION


*The intended function of this sleeve is to provide range of motion control and medial/lateral stability through compression for the treatment of mild knee instabilities. Review instructions on back page carefully before application.*

For questions, please contact your physician, the professional who fit this device or Allard USA Customer Service:


CALL: 888-678-6548      FAX: 800-289-0809  
info@allardusa.com

Allard USA, Inc. 300 Forge Way, Suite 3 Rockaway, NJ 07866-2056  
www.allardusa.com


## ROM Hinge Knee Adj Thigh



1. With thigh flap open and straps loose, pull sleeve onto leg. Make sure knee cap is centered in opening. Secure thigh flap.




2. Attach posterior calf strap, then anterior calf strap (shown).



3. Attach posterior thigh strap, then anterior thigh strap (shown). Adjust straps as needed to secure hinges in place along sides of leg.

**Note:** Your ROM hinge includes inserts for range of motion control. Select desired degree of flexion and/or extension and pop in insert as shown.



### Indications / Intended Use

Post-operative rehabilitation; Sprains and strains; Tendonitis; Arthritis; Bursitis; Genu recurvatum; Cruciate ligament, meniscal and/or collateral ligament injuries.

### Contraindications

Severe knee instabilities.

### Adjustment

This brace should only be fitted by a licensed medical professional.

### Warning


Do not overuse as this can cause skin irritation. If you feel any irritations or other problems, remove sleeve. Check skin on a daily basis.

### Wash instruction


Hand wash in cold water using mild detergent. Air dry.

**CSUS**  
by Allard USA


## ROM Hinge Knee Adj Thigh



1. With thigh flap open and straps loose, pull sleeve onto leg. Make sure knee cap is centered in opening. Secure thigh flap.




2. Attach posterior calf strap, then anterior calf strap (shown).



3. Attach posterior thigh strap, then anterior thigh strap (shown). Adjust straps as needed to secure hinges in place along sides of leg.

**Note:** Your ROM hinge includes inserts for range of motion control. Select desired degree of flexion and/or extension and pop in insert as shown.



### Indications / Intended Use

Post-operative rehabilitation; Sprains and strains; Tendonitis; Arthritis; Bursitis; Genu recurvatum; Cruciate ligament, meniscal and/or collateral ligament injuries.

### Contraindications

Severe knee instabilities.

### Adjustment

This brace should only be fitted by a licensed medical professional.

### Warning

Do not overuse as this can cause skin irritation. If you feel any irritations or other problems, remove sleeve. Check skin on a daily basis.

### Wash instruction

Hand wash in cold water using mild detergent. Air dry.

**CSUS**  
by Allard USA